

Version No.9	Hisense Ltd.	EC DECLARATION OF CONFORMITY
		COUNCIL DIRECTIVE 93/42/EEC OF 14 JUNE 1993 Concerning Medical Devices

DECLARATION OF CONFORMITY

**ACCORDING to ANNEX VI and ANNEX III of the MEDICAL DEVICE DIRECTIVE
93/42/EEC, as amended by 98/79/EC, 2000/70/EC, 2001/104/EC and 2007/47/EC**

We: HISENSE Ltd., 23 Becker St., Rishon LeZion, 75359, Israel

declare that the BABYSENSE Infant Movement Respiratory Monitor (Model CU-100/2) is in conformity with the essential requirements and provisions of Annex VI and Annex III of the Medical Devices Directive 93/42/EEC and are exclusively responsible for this declaration.

The product is classified as Class IIb, in accordance with Rule 10, Annex IX of the Directive.

Standards to which conformity is declared:


EN 60601-1:1990/AC:1994
 EN 60601-1-2:2007
 EN 60601-1-6:2007
 EN 62366:2008
 EN 980:2008
 EN 1041:2008
 EN ISO 14971:2009
 EN ISO 13485:2003

Notified Body:

LNE No. 0459
 1 rue Gaston Boissier
 75724 PARIS Cedex 15, France

Authorised Representative:

Obelis S.A.
 Boulevard Général Wahis 53
 1030 Brussels, BELGIUM


 Haim Shtalryd
 HISENSE Ltd.
 Rishon Le Zion, Israel
 01/01/2012